



GUIDELINE

Continuous Positive Airway Pressure (CPAP)

Scope (Staff):	Nursing and Medical Staff
Scope (Area):	NICU KEMH, NICU PCH, NETS WA

Child Safe Organisation Statement of Commitment

CAHS commits to being a child safe organisation by applying the National Principles for Child Safe Organisations. This is a commitment to a strong culture supported by robust policies and procedures to reduce the likelihood of harm to children and young people.

This document should be read in conjunction with this [disclaimer](#)

Contents

Aim 2

Risk..... 2

Background..... 2

Indications for CPAP failure and the need for intubation 2

CPAP Equipment Set Up 3

Application of CPAP..... 4

Nursing Care..... 7

 Observations 8

 Documentation 8

Discontinuation of CPAP 8

References 9

Appendix 1: Scaphocephaly Rating Scale..... 12

Appendix 2: Developmental Positioning 13

Appendix 3: Neoseal Application..... 15

Aim

This guideline outlines the management and use of Continuous Positive Airway Pressure (CPAP) in the neonatal unit.

Risk

Inappropriate use of CPAP and inadequate monitoring can compromise patient safety and result in adverse outcomes.

Background

Continuous Positive Airway Pressure (CPAP) is used to maintain or increase functional residual capacity of the lungs, help prevent alveolar collapse, reduce the work of breathing and improve gas exchange in infants.

CPAP can be used to treat:

- Infants with clinical signs of respiratory distress.
- In the weaning process from ventilator support.
- Splinting of upper airway in infants with obstructive apnoea.
- Infants with apnoea of prematurity.

Indications of CPAP Failure and the need for escalation of care

Sudden deterioration or prolonged apnoea requires immediate action, call for assistance ([Recognising and Responding to Clinical Deterioration](#))

- Notify the shift coordinator and medical staff if
 - There is an increase in oxygen requirements
 - Increased work of breathing
- Clinical indications to consider intubation (should be discussed with the senior registrar or consultant)
 - A rising PaCO₂ to > 60mmHg and falling pH < 7.25 despite CPAP.
 - Increasing FiO₂ and hypoxia.
 - Recurrent apnoea requiring stimulation and resuscitation.
 - Incipient collapse.
 - Administration of surfactant replacement therapy.

Transition from Intubation (mechanical ventilation) to CPAP


Please refer to [Extubation: Planned and Unplanned](#)

CPAP Equipment Set Up

Equipment

Bubble flow circuit and appropriate size Hudson CPAP prongs to achieve a snug fit.

Equipment	
Oxygen-air blender	Oxygen high flow meter
Humidifier base	Heater wire and temperature probe
IV pole to attached humidifier base	1L bag sterile water & 1L bottle sterile water
Clear combi stop for outlet port on prongs	CPAP circuit
Velcro strips	Oxygen tubing



Prong sizing

Prong Size	Recommended weight
Size 0	< 700 grams
Size 1	700-1250 grams
Size 2	1250-2000 grams
Size 3	2000-3000 grams
Size 4-5	> 3000 grams

Application of CPAP to Neonate

Prior to attaching circuit to nasal prongs

- Adjust the CPAP gauge to the pressure setting of 5-7 cm/H₂O or as ordered.
- Adjust the oxygen flow to 5-8 L/min, occlude the prongs and observe the bubble chamber for bubbles (make sure combi-stop in place). Bubbling should be constant but not excessive.
- Adjust the oxygen to meet appropriate oxygen saturation levels.

Insertion and securing nasal prongs

Nasal prongs are the preferred method of delivering CPAP. A CPAP mask may be used (after discussion with shift coordinator or CNC) where there has been trauma to the nasal septum and nares from the nasal prongs.

- Select appropriate nasal prong for size of infant.
- Use clear combi stop to seal the outlet port on the prongs.
- Comfeel is to be used to assist with providing an effective seal.
 - Comfeel is to be applied prior to the insertion of the prongs.
 - Application of nasal comfeel allows correct size nasal prong selection for weight to be used on infants despite the variance of individual nares size.
 - Also assist with reducing the movement of prongs in the nares.
- NeoSeal device can be used in conjunction with comfeel when there is an inability to maintain adequate seal. See Appendix 3

Comfeel Application Steps	Additional Information
1. Apply post admission to stabilised babies receiving CPAP therapy	
2. Using plastic templates corresponding to weight range. Cut and customise to fit infants facial features.	For best results nasal comfeel holes should be punched smaller than prong diameter.
3. Apply to a clean and dry nasal area	Consider using a barrier wipe (skin prep or convacare barrier wipe). Allow area to dry this facilitates better adhesion.
4. Align punched holes with centre of nares.	
5. Gently press comfeel onto upper lip then using finger tips press comfeel upwards following contours of nose excluding any air	
6. Finally secure nasal comfeel at the bridge of nose either by overlapping	

Comfeel Application Steps	Additional Information
edges or pinching edges together to ensure a seal.	
7. After application closely monitor for signs of prolonged blanching	(NB. Transient blanching may be observed but resolves spontaneously with return of circulation and capillary refill).
Review Comfeel at each hat release to ensure the punched holes have not enlarged allowing excessive movement.	
Nasal comfeel MUST be changed at least 24 hourly with daily hygiene regimen or as required when punched holes are no longer immobilising prongs.	
If skin integrity is compromised i.e. inflamed or moist. It may be necessary to remove nasal comfeel for 24-48 hours to enable airing and healing and revert to bonnet and ties system to secure prongs.	

CPAP Mask

CPAP pressure delivery via mask may be alternated with nasal prongs 4 hourly when there is evidence of septal redness / trauma. Discuss with CNC / Shift coordinator prior to use of mask.

- Select the appropriate size mask for the infant. The mask is to be used in conjunction with a size 2 nasal prong to provide a tight seal with the mask.
- Adjust the gas flow rate between 7-10 L/m to achieve bubbling.

NOTE: Ensure correct size prongs are used when alternating between the nasal prongs and mask.

Securing CPAP circuit

CPAP Cap System

Equipment

- CPAP Cap x1 (size appropriate).
- Long Velcro strip x2.
- Short Velcro strip x2.
- Chin strap x1.
- Self-adhesive Velcro wrap x2.
- Securing device / toggle x1

Steps	Additional Information
1. Select the appropriate CPAP Cap size. To determine the correct size - measure the circumference from the nape of the neck, across the ears to the middle of the forehead (like a turban).	E.g. Not the same as a head circumference measurement.

Steps	Additional Information
2. Place cap around the nape of the neck, over ears and across the middle of the forehead (turban style).	
3. Align the stitched marking on the cap to the tip of the left ear. Position overlap (at least 2 cm) of the cap in the middle of the forehead.	No overlap means the cap is too small - use the next size up.
4. Gather the top of cap together with a twist and use toggle creating a snug fitting cap.	
5. Place self-adhesive Velcro lengthwise around the inspiratory and expiratory corrugated CPAP tubing as close as possible to the nasal prongs connection.	See equipment set-up picture
6. Position nasal prongs in nares, ensuring a downward arch.	
7. Secure the CPAP tubings in place with the long Velcro strips.	
Note: Cap should not be tight as excessive moulding will result.	
The CPAP cap is to be change weekly or when visibly soiled.	

Bonnet with Ties

To be used if excessive moulding is observed in an infant < 30 weeks gestation and/or if nasal skin integrity is compromised.

Reassess in 72 hours and recommence CPAP Cap System if moulding/excoriation has resolved.

Equipment

- Snug fitting woollen bonnet.
- Blue ties (for securing CPAP tubings).
- Velcro (hook and fastener).
- Skin prep swab.
- Comfeel.

Steps	Additional Information
1. Apply skin prep to cheeks. Cut 2 pieces of comfeel with round corners sized to fit cheeks.	
2. Apply hook Velcro cut into an oblong to each piece of comfeel on cheeks.	

Steps	Additional Information
<p>3. Apply fastener Velcro to prongs, ensuring it is wrapped all the way around the prongs TWICE to increase the distance between the nasal septum and prongs.</p>	<p>Reduces the risk of nasal and septal trauma.</p> <p>Ensure the cut edge of the wrapped Velcro strips face away from the eyes.</p>
<p>4. Apply a well-fitting bonnet with ties. The ties are to be placed appropriately, to facilitate securing of CPAP tubing with ties in the correct position.</p>	<p>i.e. When the bonnet is in-situ, they are at the level of the tip of the ears</p>

Nursing Care

- It is preferable that all infants use the CPAP Cap System.
- Parents are encouraged to participate in CPAP cares, under direct supervision of nursing staff.
- The CPAP Cap must be released for several minutes with routine handling and cares to minimise moulding of the head and pressure areas occurring.
 - Strictly 3-4 hourly for infant's < 34 weeks and 3-6 hourly for all other infants.
- When inserting the prongs into infant's nares, position the prongs in a downward arch position but not in contact with the nasal septum, then secure them to prevent distortion of the nares and compression of the septum. Ensure there is always a minimum 2 mm gap between prongs and septum.
- Observe the bubble chamber for bubbling, this ensures positive pressure is being delivered. Bubbling assists recruitment of alveoli and gas exchange.
- Removal of nasal comfeel must only be done using an adhesive removal wipe to reduce trauma to facial skin and aid quicker removal. When removed inspect skin integrity (nares, septum and nose), document any apparent redness or ulceration.
 - Do not trim comfeel that is already on the infants' skin, remove and replace if necessary.
- Babies who are very active may cause the holes in the comfeel to rapidly enlarge and ineffectively hold the prongs in place. These babies will require a soft Velcro bolster wrapped around either side of the actual prongs to physically prevent the prongs touching the septum. Ensure that the Velcro bolster is positioned so that the cut edge is placed near the baby's cheek well away from the baby's eyes.

- To help prevent the weight of the CPAP tubing dislodging the prongs from the infant's face during weighing, repositioning or cuddles with parents - wrap a strapit around both tubes approx. 20cms above the infant's head.
- Refer to Appendix 2 for Developmental Positioning of stable and unstable infants.

Observations

All infants receiving CPAP must have hourly observations documented (Heart Rate, Respiratory Rate, Oxygen Saturations. This includes assessing for leaks and ensuring adequate bubbling.

CPAP pressure, flow and amount of oxygen being delivered must also be checked and documented hourly.

The nares and septum should be viewed hourly to allow for early identification of any pressure areas.

Documentation

Document the following on the MR489, CPAP nursing check form and/or progress notes:

- Nasal comfeel last changed.
- Condition of septum and nares.
- Evidence of head moulding. (see Appendix 1)
- Infant's position.
- CPAP effectiveness.
- Date of CPAP cap change (weekly).
- Document on MR489 chart and score the presence of scaphocephaly according to the 'Scaphocephaly Rating Scale', appendix 1. (Note: preterm infants with a history of excessive moulding may need to maintain 3-4 hourly hat release past 34 weeks CGA).

Discontinuation of CPAP

Weaning CPAP in an infant who is clinically stable

- Wean FiO₂ first then CPAP pressure.
- Cease CPAP completely without cycling when clinically stable, in CPAP of 5cmH₂O and in air.

Re-establish CPAP at previous level or consider HHF if

- Requires oxygen.
- Worsening apnoea.

- Respiratory rate consistently over 60.
- Or increased work of breathing.

Or transition to HHF at 4L/min if clinically stable at CPAP of 5cmH₂O

- If previous attempt(s) to cease CPAP unsuccessful.
- Or has persistent oxygen requirement on CPAP.

Interval between attempts to wean should be 48 hours or greater.

Related CAHS internal policies, procedures and guidelines

Neonatology Clinical Guideline

- [Extubation: Planned and Unplanned](#)
- [Monitoring and Observation Frequency](#)
- [Skin Care Guideline](#)


References

1. Blease, J. Step by Step Guide: Nursing Care Issues in CPAP. Journal on Neonatal Nursing.
2. Davis PG, Henderson-Smart DJ. Nasal continuous positive airways pressure immediately after extubation for preventing morbidity in preterm infants. The Cochrane Database of Systematic Reviews 2003, Issue 2. Art. No.: CD000143. DOI: 10.1002/14651858.CD000143.
3. Davis, P, Jankov R, Doyle L and Henschke P. 1998. Randomised controlled trial of nasal continuous positive airway pressure in the extubation of infants weighing 600g to 1250g. Archives of Diseases in Childhood Fetal and Neonatal Edition. 79 (1);F54-7.
4. De Klerk AM, De Klerk RK. Nasal continuous positive airway pressure and outcomes of preterm infants. Paediatr. Child Health (2001) 37, 161-167.
5. De Paoli AG, Davis PG, Faber B, Morley CJ. Devices and pressure sources for administration of nasal continuous positive airway pressure (NCPAP) in preterm neonates. The Cochrane Database of Systematic Reviews 2002, Issue 3. Art. No.: CD002977. DOI: 10.1002/14651858.CD002977.
6. Ho JJ, Subramaniam P, Henderson-Smart DJ, Davis PG. Continuous distending pressure for respiratory distress syndrome in preterm infants. The Cochrane Database of Systematic Reviews 2002, Issue 2. Art. No.: CD002271. DOI: 10.1002/14651858.CD002271.
7. Klausner, J, Lee A and Hutchinson, A. 1996. Decreased Imposed Work with a New Nasal Continuous Positive Airway Pressure Device. Pediatric Pulmonology 22: 188-194.
8. Narendran V, Donovan EF, Hoath S.B., Akinbi HT, Steichen JJ, Jobe AH. Early Bubble CPAP and Outcomes in ELBW Preterm Infants. Journal of Perinatology 2003; 23:195-199.
9. Robertson, NJ, McCarthy, LS, Hamilton, PA and Moss, AL. 1996. Nasal deformities resulting from flow driver continuous positive airway pressure. Archives of Diseases in Childhood Fetal and Neonatal Edition. 75(3):F209-12.
10. Subramaniam P, Henderson-Smart DJ, Davis PG. Prophylactic nasal continuous positive

airways pressure for preventing morbidity and mortality in very preterm infants. The Cochrane Database of Systematic Reviews 2005, Issue 3. Art. No.: CD001243. DOI: 10.1002/14651858.CD001243.pub2.

11. Wong, W, Fok, TF, Ng PC, Chui, KM and To KF. 1997. Vascular air embolism: A rare complication of nasal CPAP. *Journal of Paediatrics and Child Health*. 33(9):444-5
12. Blease, J. Step by Step Guide: Nursing Care Issues in CPAP. *Journal on Neonatal Nursing*.
13. Davis PG, Henderson-Smart DJ. Nasal continuous positive airways pressure immediately after extubation for preventing morbidity in preterm infants. The Cochrane Database of Systematic Reviews 2003, Issue 2. Art. No.: CD000143. DOI: 10.1002/14651858.CD000143.
14. Davis, P, Jankov R, Doyle L and Henschke P. 1998. Randomised controlled trial of nasal continuous positive airway pressure in the extubation of infants weighing 600g to 1250g. *Archives of Diseases in Childhood Fetal and Neonatal Edition*. 79 (1);F54-7.
15. De Klerk AM, De Klerk RK. Nasal continuous positive airway pressure and outcomes of preterm infants. *Paediatr. Child Health* (2001) 37, 161-167.
16. De Paoli AG, Davis PG, Faber B, Morley CJ. Devices and pressure sources for administration of nasal continuous positive airway pressure (NCPAP) in preterm neonates. The Cochrane Database of Systematic Reviews 2002, Issue 3. Art. No.: CD002977. DOI: 10.1002/14651858.CD002977.
17. Ho JJ, Subramaniam P, Henderson-Smart DJ, Davis PG. Continuous distending pressure for respiratory distress syndrome in preterm infants. The Cochrane Database of Systematic Reviews 2002, Issue 2. Art. No.: CD002271. DOI: 10.1002/14651858.CD002271.
18. Klausner, J, Lee A and Hutchinson, A. 1996. Decreased Imposed Work with a New Nasal Continuous Positive Airway Pressure Device. *Pediatric Pulmonology* 22: 188-194.
19. Narendran V, Donovan EF, Hoath S.B., Akinbi HT, Steichen JJ, Jobe AH. Early Bubble CPAP and Outcomes in ELBW Preterm Infants. *Journal of Perinatology* 2003; 23:195-199.
20. Robertson, NJ, McCarthy, LS, Hamilton, PA and Moss, AL. 1996. Nasal deformities resulting from flow driver continuous positive airway pressure. *Archives of Diseases in Childhood Fetal and Neonatal Edition*. 75(3):F209-12.
21. Subramaniam P, Henderson-Smart DJ, Davis PG. Prophylactic nasal continuous positive airways pressure for preventing morbidity and mortality in very preterm infants. The Cochrane Database of Systematic Reviews 2005, Issue 3. Art. No.: CD001243. DOI: 10.1002/14651858.CD001243.pub2.
22. Wong, W, Fok, TF, Ng PC, Chui, KM and To KF. 1997. Vascular air embolism: A rare complication of nasal CPAP. *Journal of Paediatrics and Child Health*. 33(9):444-5

This document can be made available in alternative formats on request.

Document Owner:	Neonatology		
Reviewer / Team:	Neonatal Coordinating Group		
Date First Issued:	August 2006	Last Reviewed:	October 2021
Amendment Dates:	06/12/21	Next Review Date:	26 th October 2024
Approved by:	Neonatology Coordinating Group	Date:	26 th October 2021
Endorsed by:	Neonatology Coordinating Group	Date:	
Standards Applicable:	NSQHS Standards:  Child Safe Standards: 1,10		

Printed or personally saved electronic copies of this document are considered uncontrolled



Healthy kids, healthy communities













Compassion Excellence Collaboration Accountability Equity Respect

Neonatology | Community Health | Mental Health | Perth Children's Hospital

Appendix 1: Scaphocephaly Rating Scale

Scaphocephaly Rating Scale

Pre Term infants are at risk of altered skull shape called Scaphocephaly. This is due to the altered external environment during a time of rapid brain growth.

	Nil	Minimal	Moderate	Severe	
Top View					Posterior bulging
Side View					Frontal bossing
Front View					Temporal narrowing
	<p>Nil:</p> <ul style="list-style-type: none"> • Regular position changes 	<p>Minimal:</p> <ul style="list-style-type: none"> • Regular position changes as appropriate/ tolerated 	<p>Moderate:</p> <ul style="list-style-type: none"> • Regular position changes - progress to side lying • Use peanut pillow for cares • Alert Physio • Check for head turn preference 	<p>Severe:</p> <ul style="list-style-type: none"> • Regular position changes - progress to side lying • Use peanut pillow for cares • Contact CNC to consider soft CPAP beanie • Consider using supine positioning 	

Appendix 2: Developmental Positioning for infants on CPAP

Once an infant is stable on CPAP then rotation through different positions including supine will contribute to infant's developmental care. There is no specified gestational age or weight criteria. Nutrition via continuous milk feeds (CMF) does not stop the use of a supine position. Stable is defined as:

- Stable O₂ requirements for 48 hours.
- Minimal sternal recession when in supine position.
- Respiratory Rate < 80 breaths per minute with minimum work of breathing (e.g. no increase in use of intercostals/accessory muscles and no rising CO₂).
- Occasional desaturation episodes and self-resolving bradycardias.
- No apnoeas.
- Tolerating 75% full intermittent tube feeds or CMF.
- Minimal OGT aspirates 4 hourly.
- No vomiting.
- Infant must be able to maintain appropriate temperature for gestation and weight
- Utilise position chart below to select sequence of positions. This sequence should be customised to suit each individual infant. Any position that affects the stability of infants condition should be avoided until the infant's condition is stable enough to tolerate it

Stable baby

C1 ¼ turn Supine on Rt/C2 ¼ turn supine on Lt side



Front view



Back view

D1 Supine facing Lt /D2 Supine facing Rt



CPAP babies In white cot swaddled to contain
A1 & A2 C1 & C2



Stable baby:

- Minimal respiratory assistance CPAP=5-6cm
- FiO₂ = 30%
- No need for inotropes, colloid support to maintain BP
- No sepsis or respiratory or haemodynamic instability
- Tolerating 75% full enteral feeds

Unstable baby

A1 Prone facing Lt / **A2** Prone facing Rt



Front view



Top view

B1 ¼ turn Prone facing Lt/ **B2** ¼ turn facing Rt



Front view



Back view

CPAP babies who only tolerate ¼ turn prone - when cycling off CPAP please place prone.

Check with coordinator or CNC before implementing Stable baby position changes

Appendix 3: NeoSeal Application

NeoSeal's are designed to help improve the seal and skin integrity when using bubble CPAP on neonates. The NeoSeal creates a barrier between the neonate's septum, nares and the CPAP prongs. NeoSeal's are designed to adhere to the nasal prongs, not the neonates skins to provide a soft bumper to protect the nasal septum and help to prevent skin break down NeoSeal's should be used in conjunction with comfeel and a wound care plan for septal/ nasal trauma. NeoSeal's come in three different sizes micro, small and large.

Aim

The aim of the NeoSeal is to help maintain a seal and skin integrity of the neonate when using bubble CPAP. NeoSeals maximise the CPAP effectiveness by maintaining consistent adequate PEEP thus minimising atelectasis and reduces over handling. These measures in turn aim to decrease the length of stay.

Criteria for use

Inability to maintain consistent CPAP pressure (bubbling) despite optimal patient positioning with the use of comfeel

Sepal/nasal trauma

Directions for use

Step 1. Peel NeoSeal from liner

Step 2. Slip the NeoSeal over nasal prongs with **adhesive facing the nasal prongs**

Step 3. Place nasal prongs as per CPAP guidelines

Step 4. Replace NeoSeal ever 3-5 days or when wet or soiled



Sizes

Micro: CPAP prongs size 0-1

Small: CPAP prongs size 2

Large: CPAP prongs size 3